## **AZIDE DEXTROSE BROTH**

A selective medium for the detection of enterococci.

Dehydrated media			
Code number:	500 g: ADB20500, 5 kg: ADB25000		
Colour:	Yellowish		
Appearance:	Homogeneous hygroscopic powder		
pH before autoclaving (25 °C):	7,0 – 7,4		

**Direction:** Suspend **35 g** in one litre of distilled water and heat gently to dissolve the medium completely. Dispense into final containers and sterilise by autoclaving at 115 °C for 15 minutes.

## Warning!

The medium is heat sensitive. No further sterilisation is necessary or desirable.

Prepared media		
Bottled media:	100 ml: ADB30100, 500 ml: ADB30500	
Tubed media:	150 x 15 mm: ADB40010 (10 ml)	
Colour:	Yellowish	
pH (25 °C):	7,1 – 7,3	

**Direction:** Dispense the bottled media aseptically into sterile final containers. Media in tubes are ready to use.

## FORMULA in g/l

Peptones	19,8
Glucose	7,5
Sodium chloride	7,5
Sodium azide	0,2

**Note:** The typical formula can be adjusted to obtain optimal performance.

**Storage conditions:** Store the dehydrated media tightly closed in a dry place at room temperature. Store the bottled and tubed media protected from light at room temperature. Use before the expiry date on the label.

## **Quality control:**

<b>Test strains</b>	Incubation temp: 37 °C	Growth	Incubation time: 24 h
Enterococcus faecalis	ATCC 29212	Good	
Escherichia coli	ATCC 25922	Inhibited	

References: Mallmann and Seligmann (1950) Am. J. Public Health 40: 286.

In vitro diagnostic - for professional use only!